COMPACT, HIGH-PERFORMANCE THERMOGRAPHIC CAMERA



# Ceres T 1280 Series



HIGH-RESOLUTION LWIR THERMOGRAPHIC CAMERA

### **KEY FEATURES**



COMPACT AND HIGH-RESOLUTION

SUPERIOR ON-BOARD IMAGE PROCESSING PERFORMANCE

FLEXIBLE OPTICAL MOUNT & LENS OPTIONS The Ceres T 1280 series is based upon the Dione 1280 OEM thermal imaging core with 1280x1024 pixels and 12  $\mu$ m pixel pitch.

The camera offers superior on-board thermographic performance (accuracy, stability) in the temperature range up to 400 °C.

The Ceres T 1280 camera outputs full frame images at 60 Hz via either a CameraLink or at 45 Hz via GigE Vision interface. The compact size, excellent thermographic stability and accuracy, and GenICam compliant interfacing allow for easy integration in demanding industrial thermography applications.

## **Ceres T 1280 Series**



#### **KEY PERFORMANCES**

Image format/Pixel pitch	1280 x 1024 pixels/12 μm
Integration type	Rolling Shutter
Spectral range	8 -14 μm
Max frame rate (full frame)	45 Hz (GigE); 60 Hz (CL)
Power consumption	5.5 W (GigE); 5 W (CL)
Power supply voltage	DC 12 V

#### **FUNCTIONS & INTERFACES**

Digital output format	GigE; CL
Ambient operating temperature range (*)	From -40°C to +70°C
Storage temperature range	From -40 °C to +85 °C
Detector NETD	<50 mK (at 30Hz, 300K, F/1)
Shock / Vibration	25g, 3ms, IEC 60068-2-27 / 2g, IEC 60068-2-6

(\*) Defining the limitations and restrictions of the thermographic mode (from +10°C to +50°C)

#### **PRODUCT SELECTOR GUIDE**

XEN-000739	[Ceres -	Г 1280	GigE 50	) mK (60	Hz)]
XEN-000743	[Ceres <sup>-</sup>	Т 1280	CL 50 r	nK (60 H	z)]

XEN-000740 [Ceres T 1280 GigE 50 mK (9 Hz)] XEN-000744 [Ceres T 1280 CL 50 mK (9 Hz)]

advancedimaging@exosens.com



in 🗙 f 🔚 exosens.com



© Xenics. The information furnished is believed to be accurate and reliable, but is not guaranteed and is subject to change without notice. No liability is assumed by Xenics nor by any Exosens Group companies. Performance data represents typical characteristics as individual product performance may vary. Customers should verify that they have the most current Xenics product information before placing orders. Texts and pictures may not be considered as contractually binding. This document may not be reproduced, in whole or in part, without the prior written consent of Xenics.